ANALIZA BIAŁEK

100013910

Laboratory of Analysis

Client: Nexus peptides
Order number: 100013910
Name: EPITHALON
Number of samples: 1 vial

Batch number: EPI-0426
Receive date: 2026-04-10
Start date: 2026-04-13
Enda date: 2026-04-16

Sample identification

Substance	Method of taking a sample	Form
Epithalon	Sample provided by customer	White powder in a glass vial

Results

Substance	Sample name	Content	Purity
Epithalon	EPITHALON	10,74mg	99%

Research comment

The test sample was applied in the amount of 3µl to the BioBasic 8 Thermo Scientific column. During the course of the analysis with the method a linear gradient in the solutions of H20 and Acetonitrile with the addition of 0.1% trifuluoroacetic acid, a peak of the sought substance appeared. Its content, when compared to the reference chromatogram, is 10,74mg. Height the measurement error is + - 0,05mg. The actual state of the sample sent before the test is presented on the photo below

Photo of the test sample

Password

8EK9M34T

Analiza Białek sp. z o.o.
KRS: 0000966168
NIP: 8992922970
REGON: 521751875

Headquarters and Laboratory:
ul. Duńska 9, 54-427 Wrocław
Budynek DELTA

Lab technician:
Analysis and research specialist
mgr Arkadiusz Zając

ANALIZA BIAŁEK

100013910

Chromatogram



Method

Time	H2O + 0,1% TFA	ACN + 0,1% TFA
0min	95%	5%
2min	95%	5%
12min	5%	95%
15min	5%	95%

Password

8EK9M34T

Analiza Białek sp. z o.o.
KRS: 0000966168
NIP: 8992922970
REGON: 521751875

Headquarters and Laboratory:
ul. Duńska 9, 54-427 Wrocław
Budynek DELTA

Lab technician:
Analysis and research specialist
mgr Arkadiusz Zając